

UTMB RESPIRATORY CARE SERVICES PROCEDURE - Micro Plus Spirometer Operating Procedure	Policy 7.3.21 Page 1 of 3
Micro Plus Spirometer Operating Procedure Formulated: 9/2002	Effective: 10/10/02 Revised: 08/22/23 Reviewed: 08/22/23

Micro Plus Spirometer Operating Procedure

Purpose The purpose of the Micro Spirometer Operating Procedure is to ensure the proper maintenance, tracking, and operation of the hand-held unit to provide accurate bedside pulmonary function testing.

Audience Respiratory Care Practitioners.

Scope The handheld spirometer will be utilized by trained RCS Personnel to perform bedside pulmonary functions. It is a precision instrument designed to measure Forced Expired Volume in one second, FEV₁, Forced Vital Capacity (FVC) Forced Expiratory Ratio (FER), and Peak Expiratory Flow.

Physician's Order An order for bedside pulmonary function screening by Respiratory Care Services is required unless the patient is receiving bronchodilator therapy.

Procedure

Step	Action
1	Verify Physicians order.
2	Assemble the spirometer with disposable mouthpiece.
3	Explain procedure to patient and demonstrate maneuver.
4	Switch the unit on, and follow manufacturer guidelines for setup. Instruct the patient to "Breathe in until your lungs are completely full, now seal your lips around the mouthpiece and blow out as hard and as fast as possible until you cannot push any more air out"
5	When the patient has completed this maneuver, obtain results.

Procedure Continued

Step	Action
-------------	---------------

UTMB RESPIRATORY CARE SERVICES PROCEDURE - Micro Plus Spirometer Operating Procedure		Policy 7.3.21 Page 2 of 3
Micro Plus Spirometer Operating Procedure	Formulated: 9/2002	Effective: 10/10/02 Revised: 08/22/23 Reviewed: 08/22/23

6	Perform three adequate maneuvers.
7	Record pulse oximetry values and any pertinent comments in the comments section.
8	Document results in EPIC.
9	The Hand Held Spirometer calibration is dependent only on the physical geometry of the digital volume transducer and provided that the transducer remains undamaged, will remain stable indefinitely. The unit should not therefore require recalibration.

Tracking

To be assigned each shift by the Team Leader. The therapist will return the unit to the Team Leader at the end of their shift or pass it on to the therapist assigned to that spirometer on the next shift.

If there is physical damage to the unit, return it to the Team Leader. It will be sent to Micro Medical for repair and recalibration.

Maintenance Follow manufacture guidelines.

Infection Control

Patients in isolation should be given a disposable peak flow meter. Otherwise, each patient on bronchodilator therapy should be given a disposable mouthpiece with a one-way valve (Microcheck Mouthpiece #3395) with the initial therapy.

Follow procedures as outlined Healthcare Epidemiology Policies and Procedures: #2.24 Respiratory Care Services.

<http://www.utmb.edu/policy/hcepidem/search/02-24.pdf>

References

Micro Spirometer Operating Manual

AARC Clinical Practice Guideline: Spirometry, Respiratory Care 1996; 41(7): 629-636, 1996 Update

Jonathan Dakin, Elena Kourteli, Robert Winter, Making Sense of Lung Function Tests: A Hands-On Guide Edward Arnold; (July 2003)

Scanlan, C., Sheldon, R., Spearman, C., Egan's Fundamentals of Respiratory Care, Eighth Edition, Mosby; June 2, 2003

UTMB RESPIRATORY CARE SERVICES PROCEDURE - Micro Plus Spirometer Operating Procedure	Policy 7.3.21 Page 3 of 3
Micro Plus Spirometer Operating Procedure Formulated: 9/2002	Effective: 10/10/02 Revised: 08/22/23 Reviewed: 08/22/23

Robert E. Hyatt, et al Interpretation of Pulmonary Functions Tests: A Practical Guide Lippincott Williams & Wilkins Publishers; 2nd edition (May 2003)

Robert John Hancox, Kenneth Frank Whyte Pocket Guide to Lung Function Tests McGraw-Hill Professional; 1st edition (May 1, 2002)

Dana F. Oakes, Clinical Practitioners Pocket Guide to Respiratory Care, Health Educator Publications; 5th edition June 2000

David W. Chang, Respiratory Care Calculations, Delmar Learning; 2nd edition December 17, 1998

Gregg E. Ruppel Manual of Pulmonary Function Testing Mosby; 7th edition, St. Louis: Mosby; 1998.
